

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

AUG 29 2009

DataMaster cdm S/N		130195	
Supervisor/Operator Performing the Verification Procedure:			
A	Name	Robert J. Ely	ID# 2681 Date 08-26-2009
	Agency	Craig Police Department	Phone # (907) 826-3330
	Instrument Location	506 2 <sup>nd</sup> Street, Craig Alaska 99921	
Alco S/N		Target Value	High Pressure
X172978		.079	1200 PSI
Alco Test Value Average		.084 .083	
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		<i>Robert J. Ely</i>	
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

*Nita J. Bolz*  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 17<sup>th</sup> day of Sept, 2009.

*Carolyn M. Noland*  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

- C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195

AUGUST 26, 2009

OPERATOR'S NAME:

ELV, RJ

OPERATOR'S NUMBER: 2681

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VER OF CAL

O.L. #: 123456789

DEPT/AGENCY: CR61

CASE/REPORT: 2009-XXXX

TEST TYPE: U

ALCO TARGET VALUE: .079

ALCO S/N: X172978

### — BREATH ANALYSIS —

.079 ADJUSTED FOR 30.30 in  
ALCO TARGET .000 11:29  
BLANK TEST .000 11:30  
INTERNAL STANDARD VERIFIED 11:30  
ALCO TV 30.30 in .004 11:31  
BLANK TEST .000 11:31  
SUBJECT SAMPLE .000 11:32  
BLANK TEST .000 11:33  
ALCO TV 30.30 in .003 11:33  
BLANK TEST .000 11:34

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130195

AUGUST 26, 2009  
TIME 11:27

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY  
PROGRAM: OKAY  
SOFTWARE DATE: 02/20/01  
HEATERS  
SAMPLE CHAMBER: 49c  
BREATH TUBE: 42c  
BAROMETER: 30.30 in  
FLOW DETECTOR: OKAY  
PUMP  
HIGH SPEED: OKAY  
DETECTOR: OKAY  
FILTERS: OKAY  
QUARTZ STANDARD: OKAY  
CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHI  
HIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~